

EC Declaration of Conformity

Certificate No.: ML20240704001

Manufacturer:
Microlife Corporation
9F, 431, RuiGuang Road, NeiHu,
Taipei, 11492, Taiwan, R.O.C.

whose single Authorized Representative:
Microlife UAB
P. Lukšio g. 32, 08222 Vilnius, Lithuania

We, the manufacturer, herewith declare that the products/product category
Automatic-inflation electronic GMDN 45617
shpygmomanometer, portable, arm/wrist
Atrial fibrillation-detection blood pressure monitor GMDN 64511
Blood pressure long-term ambulatory recorder GMDN 36888
Please refer to the Appendix for model list.

meet the provisions of Directive 93/42/EEC amended by 2007/47/EC and Regulation (EU) 2017/745 Article 120(3), which apply to them.

The medical devices have been assigned to class IIa according to Annex IX rule 10 of the Directive 93/42/EEC. It bears the mark



The products concerned have been designed and manufactured under a quality management system according to Annex II of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body


TÜV NORD CERT GmbH
Langemarckstraße 20 45141 Essen
Certificate No.: 04 232 950010

Validity from: 2019-10-22 Expiry date: 2028-12-31

following the procedure relating to the EC Declaration of Conformity set out in Annex II of Directive 93/42/EEC, and in conformity to the following standards or other normative documents:

EN IEC 80601-2-30:2019 (IEC 80601-2-30:2018)
EN 60601-1:2006+A1:2013+AC: 2014 (IEC 60601-1:2005+CORR.1:2006+CORR.2:2007+AM1:2012(or IEC60601-1:2012))
EN 60601-1-2:2015 (IEC 60601-1-2:2014)
EN 60601-1-11:2015+A1:2021(IEC 60601-1-11:2015+AMD1:2020)
EN 62304:2006+A1:2015 (IEC 62304:2006+AMD1:2015)
EN 60601-1-6:2010+A1:2015+A2:2021(IEC 60601-1-6:2010+AMD1:2013+AMD2:2020)
EN 62366-1:2015+AC:2015+AC:2016+A1:2020(IEC 62366-1:2015+AMD1:2020)
EN ISO 10993-1:2020(ISO 10993-1:2018)
EN ISO 10993-5: 2009(ISO 10993-5: 2009)
EN ISO 10993-10:2023(ISO 10993-10:2021)
EN ISO 10993-12:2021(ISO 10993-12:2021)
EN ISO 14971:2019+A11:2021(ISO 14971:2019)
EN ISO 15223-1:2021(ISO 15223-1:2021)
EN1041:2008+A1:2013
EN ISO 81060-2:2019(ISO 81060-2:2018)

The above mentioned declaration of conformity is exclusively under the responsibility of Microlife Corporation, and this declaration is valid until Dec 31, 2028.


Ray Lei, Product Manager, Issued Date


Ariel Wang, Global RA & QM Director, Issued Date

Reference or article number	Class	GMDN code
BP3MX1-3C(WatchBP Home A BT)	Ila	64511
BP3MX1-3 (WatchBP Home A)	Ila	64511
BP3MX1-1 (WatchBP Home)	Ila	45617
BP3MX1-5 (WatchBP Home S)	Ila	64511
BP3MX1-4 (WatchBP Home N)	Ila	64511
TWIN200 ABI (WatchBP Office ABI)	Ila	64511
TWIN200 AFS (WatchBP Office AFIB)	Ila	64511
TWIN200 VSR (WatchBP Office Vascular)	Ila	64511
BP3MZ1-1A (WatchBP O3 AFIB 1G)	Ila	36888
BP3MZ1-1 (WatchBP O3 1G)	Ila	36888
BP3SK1-3B (WatchBP Office 2G)	Ila	45617
BP3SZ1-1 (WatchBP O3 2G)	Ila	36888